FUTEK MODEL QLA401 Custom Autoclavable Thru Hole Load Cell ITEM NUMBER: N/A INCH [mm] R.O.= Rated Output Ø0.122 ± 0.002 Ø0.420 $[3.1 \pm 0.1]$ WIRING CODE [10.7] + Signal + Excitation - Excitation - Signal Ground Ø0.240 RED **BLACK GREEN** WHITE To Body [6.1] PIN C PIN D PIN A PIN B PIN E +Output (Compression) 0.31 LOADING AREA \emptyset 0.530 ± 0.002 [7.9] $[13.5 \pm 0.05]$ 0.045 [1.1] Ø0.030 0.132 0.127 [0.8] [3.4] [3.2] B ¢ D 4X 0 050 0.058 [1.3] [1.5] [1.3] 4X Ø0.015 [0.4] DO NOT CONTACT

SPECIFICATIONS:

RATED OUTPUT 1.5 mV/V nom. CAPACITY 175 lb. [777N] SAFE OVERLOAD 150% of R.O. 0.45 to 0.95 mV/V ZERO BALANCE EXCITATION (VDC OR VAC) 10 MAX BRIDGE RESISTANCE 1000Ω nom.

NONLINEARITY ±1% of R.O. HYSTERESIS ±1% of R.O. NONREPEATABILITY ±0.25% of R.O. TEMP. SHIFT ZERO ±0.02% of R.O./°F (±0.036 of R.O./°C)

TEMP. SHIFT SPAN ±0.02% of LOAD/°F (±0.036 of LOAD/°C) COMPENSATED TEMP. 59 to 104 °F (15 to 40 °C)

59 to 104 °F (15 to 40 °C) OPERATING TEMP 32 to 279 °F (0 to 137 °C) NON OPERATING TEMP

INCONEL MATERIAL DEFLECTION 0.0005 [0.013] nom.

PIN HEADER: PLATING- GOLD 0.00005" OVER NICKEL 0.0002", Ground Pin NICKEL ONLY

PRESSURE 34 PSI MAX

CAPABLE OF SURVIVING HUNDREDS OF STERILIZATION CYCLES STERILIZATION

CE RoHS 2011/65/EU

PROPRIETARY DESIGN

MOUNTING SURFACE

Ø0.480

[12.2]

CUSTOMER APPROVAL- COMPANY:	OUTLINE DRAW	NG	This drawing is submitted solely for the information and exclusive use of the original addressee. It is not to be
CUSTOMER APPROVAL- NAME / DATE:	STANDARD NOTES: (Unless Otherwise Specified) ALL DIMENSIONS ARE IN INCHES	ANOLE	divulged in whole or in part, by any firm or individual without written permission from:
REVISIONS: (Refer to dwg # revision sheet)	DRAWING INTERPRETATION DIMS. PER ASME-Y14.5M REMOVE BURRS AND BREAK SHARP EDGES .005015	ANGLE: ± 1/2*	MODEL: QLA401
	THREADS PER HANDBOOK H-28 DIMENSIONS ARE SHOWN AFTER PLATING 2rd ANGLE DROLL	CHAMFER: ± 5°	DRAWN BY: M.Lisiak

TOLERANCE: .X ± 0.1" .XX ± 0.01" .XXX ± 0.005

3rd ANGLE PROJ.

± 5°

DWG No.: CREATED DATE: 2/10/2017 UKAWNBY: M.Lisiak APPROVED BY: R. Walker APPROVED DATE: |0/|3/20| CHECKED BY: CAGE: 1X8M6 SHEET: 1 OF 1

ADVANCED SENSOR TECHNOLOGY, INC